

Leak Test Certification

Atlas Technologies does hereby certify that this part or assembly passes the standard leak rate test of 1×10^{-9} Torr Lt. He/ Sec. or better. Leak check was performed using a Pfeiffer Vacuum QualyTest Dry HLT 270 helium leak detector.

Customer: Argonne National Laboratory
P/N: ATLAS CUSTOM - LCLS (#33)

Project #: _____

Tested by: Robin S.

Date: 12-13-07

Leak Test Certification

Atlas Technologies does hereby certify that this part or assembly passes the standard leak rate test of 1×10^{-9} Torr Lt. He/ Sec. or better. Leak check was performed using a Pfeiffer Vacuum QualyTest Dry HLT 270 helium leak detector.

Customer: Argonne National Laboratory
P/N: ATLAS custom - LCLS (# 33)

Project #: _____

Tested by: Robin S.

Date: 12-13-07